

SPRING 5UV

Cat. no: 5DS-TOC-UV





Technical specification:

- The device works under tap water pressure.
- Water purification levels:
 - sediment filtration – prefilter 5 μ m,
 - module A (sediment-carbon-softening),
 - demineralization on a spectrally clean mixed ion exchange resin,
 - UV lamp 254 nm,
 - cascade microfiltration capsule 0,45/0,2 μ m.
- Automatic and unattended system operation.
- Equipped with a pump increasing feed water pressure.
- Two independent water intake points:
 1. Second purity class (PN-EN ISO 3696:1999, ASTM, CLSI) - nozzle reach min. 2 m, equipped with a pressure tank.
 2. First purity class (PN-EN ISO 3696:1999) - with a 0,2 μ m microfiltration capsule.
- System equipped with a 10 dm³ pressure storage tank.
- Optional replacement with a bigger tank (40dm³, 80dm³ and more).
- The possibility of installing additional water intake point for general-purpose water (third class (PN-EN ISO 3696:1999)).
- Automated system shutdown when the tank is full.
- Optional connection to an autoclave, washer machine etc.
- User-performed maintenance procedures (easy disposables replacement).
- Intended to be fed by cold water: 5-40°C.
- Energy consumption less than 70W.
- Optional user-performed device installation.
- Power supply: 220-240V/50Hz.

Dimensions (SxGxW): 235x440x510 mm

Tank 10l: height: 390 mm, diameter: 250 mm

Functions monitoring the device:

- The device is equipped with a 24V automatics with a microprocessor control and measurement system, that includes:
 - color display screen with a Touch Panel,
 - conductometer measuring conductivity and temperature of feed water, after reverse osmosis and purified water (measured in μ S/cm or M Ω m),
 - clock displaying date and time,
 - information about current system status,
 - information about the membrane module retention level,
 - alarm informing about necessity to replace module A,
 - alarm informing about necessity to replace ionex resins,
 - alarm informing about necessity to replace UV lamp radiator,
 - alarm informing about necessity to replace microfiltration capsule,
 - graphic and sound alarm signal,
 - maintenance dates preview,
 - tank fill level,
 - built-in RS 232 connection to personal computers allowing to adjust maintenance frequency and alarm levels,
 - built-in USB connection to personal computers allowing to adjust maintenance frequency and alarm levels.
- Software.
- Built-in manometer measuring feed water pressure.

Functions protecting the device:

- Pump shutdown when:
 - the feed water pressure is too low (lack of feed water) - low pressure sensor,
 - the tank is full - high pressure sensor.
- Thermal protection of the RO module, automated system shutdown when the feed water temperature is below 4°C or above 40°C.
- Can be automatically shut down when any alarm occurs.
- System autostart.
- Notification/alarm preview.

**Standard:**

Water purified by the Spring device fits the requirements of the ISO 3696:1999, ASTM, CLSI standard for I* and II purity class, microbiological and physicochemical parameters match the FP requirements for purified production water

Application:

Obtained water may be used for instrumental analyses AAS, ICP/MS, IC*, HPLC*, GC, bacteria cultures*, biochemical analyses*.

*point with a 0,2µm microfiltration capsule

Required connections:

- cold water connection ½" or ¾",
- 220-240V socket,
- drain.

General information:

- fed by: tap water
- efficiency: min. 5 l/h
- conductivity: 0,06 µS/cm
- resistivity: 18,2 MOhm*cm
- bacteria < 1 cfu/ml*
- particles > 0,2 µm < 1/ml*

model	Sediment prefilter 5µm	Module A2	Modules 2x H7TOC	MF capsule 0,2µm	UV Lamp radiator 254nm
SPRING 5UV	+	+	+	+	+
Lifetime	6 months*	6 months*	2x 2000 dm ³ **	12 months	8500 h
Cataloge no.	EO-005-10	EO-MA-12	EJ-2000-0	EM-SP-20	EUV-254-HLP

* The life of a filter cartridge can be affected by the flow, it's characteristic as well as the level and type of the contamination.

** * Volume of the purified water depends on the quality of the feed water, the maximum amount of the dissolved salt in the feed water - 1200 mg/l.